

Test Report

Date: 11th Jul. 2018

Client name: WACKER CHEMICALS (CHINA) CO., LTD

Client address: Bldg. 3, 1535 Hongmei Road Caohejing Hi-Tech Park Shanghai 200233, China

Assignment ID: 14A1802067

Sample No.: 14S18006091-01

Report on the submitted sample identified by the client as below:

Product Name	ELASTOSIL®LR 3038/50 K1 CN
Quantity Received	1 bag
Batch number	ZR13214
Expiry date	unlimited storage life
Type of material	Synthetic Elastomer
Sample Receiving Condition	Room temperature
Sample Receiving Date	13 th Apr.2018
Testing Period	15 th May.2018 –18 th May.2018

Test Requested, Test Method and Test Results:

Please refer to the following page(s), **Attachment 1**.

The test was carried out by SGS subcontractor certified ISO 17025 by CNAS. The results contained in this Report are in the scope of ISO 17025 certification.

Signed for and on behalf of SGS

.....*Sia Tong*.....
Sia Tong
Life Science Quality Assurance
Authorized Signature



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, identification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing/ inspection report & certificate, please contact us at telephone: (86-755)83071443 or email: CN.Doccheck@sgs.com

3rd Building No.889 Yishan Road Xuhui District, Shanghai, China 200233 t:(86-21)61152197 f:(86-21)64951517 www.sgs.com.cn
中国,上海,徐汇区宜山路 889 号 3 号楼 邮编: 200233 t:(86-21)61152197 f:(86-21)64951517 e: sgs.china@sgs.com

SHLS 225594

Client name: WACKER CHEMICALS (CHINA) CO., LTD
Client address: Bldg. 3, 1535 Hongmei Road Caohejing Hi-Tech Park Shanghai 200233, China
Assignment ID: 14A1802067
Sample No.: 14S18006091-01

Attachment 1: Test for Pyrogenicity

SUMMARY

The test article: ELASTOSIL®LR 3038/50 K1 CN was extracted in 0.9% sterile nonpyrogenic sodium chloride injection (SC). The extract was evaluated in the rabbit for material medicated pyrogenicity. This study was conducted based on the requirements for ISO 10993-11:2006: Biological Evaluation of Medical Device, Part 11: Test for Systemic Toxicity; ISO 10993-12: 2012: Biological evaluation of medical devices Part 12: Sample preparation and reference materials.

A single dose of 10ml/kg was injected intravenously via the marginal ear vein to each of three rabbits. Rectal temperatures were measured and recorded prior to injection and at 30 min intervals between 1 and 3h after injection. The maximum temperature rise (as compared to baseline) for each rabbit was determined, and whether the extract of the test article induced a pyrogenic response was evaluated.

Under the conditions of this study, the total rise of rabbit temperature during the 3h observation period was within acceptable limits. The extract of the test article was judged as nonpyrogenic.

MATERIALS

The test article provided by the sponsor was identified and handled as follows:

Test Article: ELASTOSIL®LR 3038/50 K1 CN
Sterilization Status : Non-sterile
Storage Conditions: Room temperature
Extract Vehicle: 0.9% sterile nonpyrogenic sodium chloride injection (SC)
Test Article Preparation: According the requirement of the sponsor, the test articles were sterilized by ethylene oxide two weeks before the treatment.

Based on the ISO10993-12:2012, the ratio of 1.25cm² : 1 ml [Surface area of the test article to volume of extraction vehicle], 125cm² of the test article was covered with 100 ml extraction vehicle under aseptic conditions for preparing the SC test extract at 37 °C for 72 hours respectively.

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, identification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing/ Inspection report & certificate, please contact us at telephone: (86-755)83071443 or email: CN.Doccheck@sgs.com

SHLS 225595

3rd Building No.889 Yishan Road Xuhui District, Shanghai, China 200233 t:(86-21)61152197 f:(86-21)64951517 www.sgs.com
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233 t:(86-21)61152197 f:(86-21)64951517 e: sgs.china@sgs.com

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

Client name: WACKER CHEMICALS (CHINA) CO., LTD
Client address: Bldg. 3, 1535 Hongmei Road Caohejing Hi-Tech Park Shanghai 200233, China
Assignment ID: 14A1802067
Sample No.: 14S18006091-01

The fresh extracts was warmed in a 38 °C water bath for a minimum of 10 min prior to injection.

Instrument Preparation: Render the syringes, needles, and glassware free from pyrogens by heating at 180 °C for 2h.

Condition of Extracts: The extract of the test articles was clear and without any special treatments.

METHODS

Test System

Species: Rabbit
Grade: New Zealand White
Source: SHANGHAI SONGLIAN LAB ANIMAL-FIELD
Sex: Male
Body Weight Range: 2.2 kg ~ 2.3kg
Age: Adult
Number of animals: Three

Animal Management:

Husbandry: Conditions conformed to “Laboratory animal-Requirements of environment and housing facilities”; “ISO 10993-2:2006: Biological evaluation of medical devices Part 2: Animal welfare requirements”.

Food: Diet was provided from Shanghai Pu Lu Teng Biological Technology Co., Ltd.

Housing: Healthy animals were acclimatized to the laboratory conditions for 7 days before the treatment, and then they were individually housed in stainless steel suspended cages identified by a card indicating the Identification No of the test article and first treatment date.

Environmental: The room temperature and humidity were monitored daily. The room temperature was 21 °C. The room humidity was 60%.

Personnel Associates involved were appropriately qualified and trained.

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, identification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755)83071443 or email: CN.Doccheck@sgs.com

SHLS 225596

3rd Building No.889 Yishan Road Xuhui District, Shanghai, China 200233 t:(86-21)61152197 f:(86-21)64951517 www.sgs.com
中国·上海·徐汇区宜山路 889 号 3 号楼 邮编: 200233 t:(86-21)61152197 f:(86-21)64951517 e: sgs.china@sgs.com

Client name: WACKER CHEMICALS (CHINA) CO., LTD
 Client address: Bldg. 3, 1535 Hongmei Road Caohejing Hi-Tech Park Shanghai 200233, China
 Assignment ID: 14A1802067
 Sample No.: 14S18006091-01

Selection: Only healthy animals were selected. Prior to using, the animals in this test were conditioned for 5d by a sham test that includes all of the steps as directed under procedure except injection.

Experimental Procedure:

Not more than 30 min prior to injection, the temperature of each animal was measured by ZRY-2D Pyrogen Tester. This temperature was recorded as the baseline temperature for the study.

Each rabbit was received a single intravenous injection of the extract via the marginal ear vein at 10ml/kg of the body weight. Rectal temperatures were measured and recorded at 30 min intervals between 1h and 3h after injection.

The maximum temperature rise (as compared to baseline) for each rabbit was determined.

Consider any temperature decreases as zero rises. If no rabbit showed an individual temperature rise of 0.5 °C or more above the baseline temperature, the extract of the test article met the requirements for the absence of pyrogens. If any rabbit showed an individual temperature rise of 0.5 °C or more, a retest was conducted using five other rabbits. If not more than three of the eight rabbits showed individual rises in temperature of 0.5 °C or more, and if the sum of the eight individual maximum temperature rises did not exceed 3.3 °C, the test article met the requirements for the absence of pyrogens.

RESULTS

No single rabbit showed a temperature increased of 0.5 °C or more above its baseline temperature.

Individual temperatures were presented below:

No	Weight (kg)	Sex	Dose volume (ml)	Baseline (°C)	Hours after injection						Maximum Rise
					0.5	1.0	1.5	2.0	2.5	3.0	
1	2.1	♂	23	39.1	39.1	39.2	39.2	39.2	39.1	39.1	0.1
2	2.1	♂	22	38.6	38.6	38.7	38.7	38.7	38.7	38.7	0.1
3	2.2	♂	23	39.1	39.1	39.0	39.1	39.1	39.1	39.1	0.0

CONCLUSION

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, identification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing/ inspection report & certificate, please contact us at telephone: (86-755)8307 1443 or email: CN.Doccheck@sgs.com

SHLS 225597


3rd Building No.889 Yishan Road Xuhui District, Shanghai, China 200233 t:(86-21)61152197 f:(86-21)64951517 www.sgs.com
 中国·上海·徐汇区宜山路 889 号 3 号楼 邮编: 200233 t:(86-21)61152197 f:(86-21)64951517 e sgs.china@sgs.com

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: CN.Doccheck@sgs.com

Client name: WACKER CHEMICALS (CHINA) CO., LTD
Client address: Bldg. 3, 1535 Hongmei Road Caohejing Hi-Tech Park Shanghai 200233, China
Assignment ID: 14A1802067
Sample No.: 14S18006091-01

Under the conditions of this study, the total rise of rabbit temperature during the 3h observation period was within acceptable limits (<0.5°C). No evidence of material-mediated pyrogenicity was observed. The extract of the test article was judged as nonpyrogenic.

PHOTOGRAPH OF THE TEST ARTICLE



Remark: Results and conclusions apply only to the test article sample tested provided by Client. Therefore, this Report contains the results obtained in the test of the provided samples only and do not express any opinion upon the lot from which the samples were drawn or any similar samples.

***End of Report ***

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, identification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



Attention: To check the authenticity of testing/ inspection report & certificate, please contact us at telephone: (86-755)83071443 or email: CN.Doccheck@sgs.com

SHLS 225598

3rd Building No.889 Yishan Road Xuhui District, Shanghai, China 200233 t(86-21)61152197 f(86-21)64951517 www.sqsgroup.com.cn
中国·上海·徐汇区宜山路 889 号 3 号楼 邮编: 200233 t(86-21)61152197 f(86-21)64951517 e sgs.china@sgs.com